
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000447
9:34 15/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:08
STERILIZING 4:24
136.4 C 224kPa 4:24
136.2 C 222kPa 4:54
136.2 C 221kPa 5:24
136.3 C 224kPa 5:54
136.4 C 224kPa 6:24
136.3 C 223kPa 6:54
136.1 C 221kPa 7:24
136.4 C 224kPa 7:54
136.4 C 224kPa 7:54
VENTING 7:54
AIR DRYING 8:51
CYCLE COMPLETE 8:57

Digital Signature #
B06628B36F181934
