

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2804

Name: RITTER2  
Sterilizer ID: V2313228  
08/28/2023 12:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.8 F 0.0  
2:00 167.2 F 0.9  
4:00 186.3 F 4.1  
6:00 206.0 F 8.1  
8:00 227.9 F 12.9  
10:00 230.8 F 8.3  
12:00 241.9 F 13.0  
14:00 240.9 F 11.5  
16:00 239.1 F 10.3  
18:00 252.0 F 16.6  
20:00 262.6 F 22.9  
22:00 272.4 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.1 F 29.2  
2:00 272.9 F 29.5  
3:00 272.5 F 29.8  
4:00 271.4 F 29.3

Min 271.4 F 29.1  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.0 F 0.0  
5:00 222.8 F 0.0  
10:00 207.5 F 0.0  
15:00 182.2 F 0.0  
20:00 177.8 F 0.0  
25:00 181.3 F 0.0  
30:00 175.2 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 23:07  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:02:05  
CYCLE COMPLETE

08/28/2023 01:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_