
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001102
13:42 01/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:56
STERILIZING 4:27
136.4 C 224kPa 4:27
136.1 C 222kPa 4:57
136.3 C 223kPa 5:27
136.3 C 224kPa 5:57
136.3 C 224kPa 6:27
136.1 C 222kPa 6:57
136.3 C 225kPa 7:27
136.4 C 225kPa 7:57
136.4 C 225kPa 7:57
VENTING 7:57
AIR DRYING 8:49
CYCLE COMPLETE 18:59

Digital Signature #
CB1FE476E4E82F9B
