

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1382

Name: RITTER2  
Sterilizer ID: V2313228  
08/09/2022 09:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.2 F 0.0  
2:00 188.2 F 0.7  
4:00 190.4 F 5.3  
6:00 210.4 F 10.7  
8:00 223.4 F 7.0  
10:00 237.1 F 13.2  
12:00 238.5 F 11.3  
14:00 238.6 F 10.3  
16:00 252.5 F 17.3  
18:00 264.1 F 24.3  
20:00 271.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.1 F 29.9  
2:00 272.8 F 30.1  
3:00 272.7 F 29.7  
4:00 272.1 F 29.4

Min 271.4 F 29.0  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 225.4 F 0.0  
10:00 233.2 F 0.0  
15:00 232.1 F 0.0  
20:00 221.5 F 0.0  
25:00 219.8 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 20:08  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:33  
CYCLE COMPLETE

08/09/2022 10:32 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_