
Midmark M11 - v1.0.0
Total Cycles: 1959

Name: RITTER 1
Sterilizer ID: V2178198
05/26/2022 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.2 F 0.0
2:00 157.9 F 3.3
4:00 179.4 F 7.5
6:00 206.2 F 13.0
8:00 223.8 F 7.3
10:00 235.8 F 12.5
12:00 235.5 F 9.3
14:00 238.3 F 9.1
16:00 247.5 F 14.1
18:00 260.1 F 21.0
20:00 271.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.4
1:00 272.6 F 29.2
2:00 271.7 F 28.5
3:00 271.9 F 29.2
4:00 272.7 F 29.2

Min 270.9 F 28.3
Max 273.2 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.2 F 0.0
5:00 227.8 F 0.0
10:00 220.4 F 0.0
15:00 226.9 F 0.0
20:00 226.8 F 0.0
25:00 223.9 F 0.0
30:00 219.0 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:25
CYCLE COMPLETE

05/26/2022 02:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____