
Midmark M11 - v1.0.5
Total Cycles: 77

Name: RITTER2
Sterilizer ID: V2313228
06/08/2021 11:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.7 F 0.0
2:00 132.0 F 1.3
4:00 138.3 F 4.4
6:00 167.1 F 8.0
8:00 190.9 F 12.2
10:00 206.3 F 5.6
12:00 220.6 F 10.3
14:00 226.3 F 8.2
16:00 237.2 F 12.9
18:00 238.8 F 11.7
20:00 251.4 F 18.0
22:00 262.8 F 24.6
24:00 272.1 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 30.0
1:00 272.7 F 30.9
2:00 272.2 F 30.3
3:00 271.5 F 30.0
4:00 272.1 F 30.9

Min 271.3 F 29.9
Max 273.0 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.2 F 0.0
5:00 239.3 F 0.0
10:00 214.0 F 0.0
15:00 200.9 F 0.0
20:00 194.5 F 0.0
25:00 190.7 F 0.0
30:00 188.8 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:14
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

06/08/2021 12:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____