

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3056

Name: RITTER 1  
Sterilizer ID: V2178198  
03/30/2023 08:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.2 F 0.0  
2:00 170.5 F 1.8  
4:00 181.3 F 5.5  
6:00 200.6 F 10.0  
8:00 220.3 F 5.9  
10:00 228.8 F 9.2  
12:00 238.7 F 9.5  
14:00 241.5 F 11.9  
16:00 239.7 F 10.1  
18:00 252.9 F 16.8  
20:00 264.6 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.4 F 29.0  
2:00 271.2 F 28.4  
3:00 272.3 F 29.2  
4:00 271.7 F 28.5  
  
Min 271.2 F 28.4  
Max 272.4 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.3 F 0.0  
5:00 225.4 F 0.0  
10:00 233.5 F 0.0  
15:00 232.7 F 0.0  
20:00 229.6 F 0.0  
25:00 229.7 F 0.0  
30:00 228.5 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 21:56  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:58  
CYCLE COMPLETE

03/30/2023 09:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_