
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001092
13:36 25/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:50
STERILIZING 3:58
136.0 C 220kPa 3:58
136.0 C 222kPa 4:28
136.4 C 225kPa 4:58
136.4 C 225kPa 5:28
136.2 C 222kPa 5:58
136.3 C 224kPa 6:28
136.3 C 224kPa 6:58
136.4 C 225kPa 7:28
136.4 C 225kPa 7:28
VENTING 7:28
AIR DRYING 8:19
CYCLE COMPLETE 17:51

Digital Signature #
B4B0BA0BF7099186
