
Midmark M11 - v1.0.5
Total Cycles: 4127

Name: RITTER2
Sterilizer ID: V2313228
11/15/2024 03:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.2 F 0.0
2:00 109.1 F 0.0
4:00 121.8 F 0.1
6:00 157.5 F 1.6
8:00 184.5 F 5.0
10:00 206.5 F 8.9
12:00 218.3 F 13.3
14:00 226.1 F 8.5
16:00 238.9 F 12.7
18:00 239.0 F 11.4
20:00 238.2 F 10.2
22:00 251.7 F 16.9
24:00 262.9 F 23.5
26:00 272.5 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.2 F 29.6
2:00 272.0 F 30.0
3:00 271.4 F 29.2
4:00 272.5 F 29.8

Min 271.2 F 29.2
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 212.2 F 0.0
10:00 203.3 F 0.0
15:00 192.3 F 0.0
20:00 187.4 F 0.0
25:00 182.1 F 0.0
30:00 178.3 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 26:15
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:04:46
CYCLE COMPLETE

11/15/2024 04:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____