
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076

Unit # : 000

WATER QUALITY

2.8uS / 1.7ppm

CYCLE NUMBER 000044

8:35 24/08/2020

RUBBER AND PLASTICS

121 C for 15 min.

CYCLE START 0:00

CONDITIONING 2:55

PRESSURIZING 7:50

STERILIZING 8:24

123.0 C 118kPa 8:24

123.4 C 120kPa 8:54

123.4 C 120kPa 9:24

123.4 C 120kPa 9:54

123.4 C 120kPa 10:24

123.4 C 120kPa 10:54

123.5 C 120kPa 11:24

123.4 C 120kPa 11:54

123.1 C 118kPa 12:24

123.0 C 120kPa 12:54

123.2 C 120kPa 13:24

123.4 C 120kPa 13:54

123.4 C 120kPa 14:24

123.1 C 118kPa 14:54

123.4 C 120kPa 15:24

122.9 C 118kPa 15:54

123.0 C 119kPa 16:24

123.1 C 118kPa 16:54

123.3 C 119kPa 17:24

123.4 C 120kPa 17:54

123.5 C 120kPa 18:24

123.4 C 120kPa 18:55

123.3 C 119kPa 19:25

123.4 C 120kPa 19:55

123.5 C 120kPa 20:25

123.4 C 120kPa 20:55

123.5 C 120kPa 21:25

123.4 C 118kPa 21:55

123.4 C 118kPa 22:25

123.4 C 120kPa 22:55

123.0 C 119kPa 23:24

VENTING 23:24

AIR DRYING 24:11
CYCLE COMPLETE 54:12

Digital Signature #
4AE8C08D53BD15D1
