
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003562
14:44 28/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:29
STERILIZING 3:26
136.5 C 222kPa 3:26
136.3 C 224kPa 3:56
136.1 C 221kPa 4:26
136.4 C 225kPa 4:56
136.4 C 226kPa 5:26
136.4 C 226kPa 5:56
136.2 C 220kPa 6:26
136.4 C 222kPa 6:56
136.4 C 222kPa 6:56
VENTING 6:56
AIR DRYING 8:37
CYCLE COMPLETE 9:14

Digital Signature #
1441DEA2CB7B3518
