

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4298

Name: RITTER 1  
Sterilizer ID: V2178198  
02/13/2024 12:33 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 102.4 F 0.0  
2:00 127.9 F 0.7  
4:00 127.7 F 2.6  
6:00 138.8 F 5.6  
8:00 162.3 F 9.2  
10:00 185.5 F 13.4  
12:00 207.7 F 5.1  
14:00 221.2 F 9.1  
16:00 232.0 F 13.5  
18:00 235.6 F 9.6  
20:00 238.2 F 9.7  
22:00 245.3 F 13.1  
24:00 256.8 F 19.1  
26:00 267.1 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.6 F 29.3  
2:00 271.8 F 28.6  
3:00 271.4 F 28.8  
4:00 272.6 F 29.2

Min 271.0 F 28.4  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 226.0 F 0.0  
10:00 227.9 F 0.0  
15:00 229.8 F 0.0  
20:00 227.6 F 0.0  
25:00 230.1 F 0.0  
30:00 227.4 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 27:34  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:06:06  
CYCLE COMPLETE

02/13/2024 01:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_