
Midmark M11 - v1.0.0
Total Cycles: 2964

Name: RITTER 1
Sterilizer ID: V2178198
03/01/2023 03:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.0 F 0.0
2:00 151.1 F 1.4
4:00 170.2 F 4.5
6:00 196.6 F 9.0
8:00 226.0 F 8.6
10:00 226.6 F 8.9
12:00 239.0 F 10.3
14:00 239.7 F 11.4
16:00 238.2 F 9.7
18:00 252.2 F 16.5
20:00 264.1 F 23.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.5 F 29.1
2:00 271.5 F 28.3
3:00 272.6 F 29.2
4:00 272.1 F 28.7

Min 271.5 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 222.3 F 0.0
10:00 233.1 F 0.0
15:00 236.2 F 0.0
20:00 227.3 F 0.0
25:00 227.7 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 21:57
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:07
CYCLE COMPLETE

03/01/2023 04:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____