
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000195
9:52 07/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 3:34
STERILIZING 4:40
136.0 C 220kPa 4:40
136.3 C 222kPa 5:10
136.4 C 224kPa 5:40
136.4 C 224kPa 6:10
136.4 C 224kPa 6:40
136.2 C 220kPa 7:10
136.4 C 224kPa 7:40
136.4 C 222kPa 8:10
136.4 C 222kPa 8:10
VENTING 8:10
AIR DRYING 9:03
CYCLE COMPLETE 9:06

Digital Signature #
8ACA06B68D0B9A53
