
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002760
14:01 30/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:19
PRESSURIZING 3:20
STERILIZING 4:26
136.0 C 220kPa 4:26
136.3 C 225kPa 4:56
136.2 C 222kPa 5:26
136.1 C 223kPa 5:56
136.3 C 224kPa 6:26
136.4 C 223kPa 6:56
136.2 C 222kPa 7:26
136.2 C 223kPa 7:56
136.2 C 223kPa 7:56
VENTING 7:56
AIR DRYING 8:50
CYCLE COMPLETE 8:58

Digital Signature #
44FE1F03D03DB475
