

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4053

Name: RITTER2  
Sterilizer ID: V2313228  
10/24/2024 01:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.3 F 0.0  
2:00 134.6 F 0.1  
4:00 142.7 F 1.6  
6:00 171.0 F 5.0  
8:00 188.8 F 8.6  
10:00 209.1 F 12.6  
12:00 220.4 F 7.1  
14:00 229.6 F 11.7  
16:00 232.7 F 8.7  
18:00 239.9 F 11.3  
20:00 244.4 F 13.1  
22:00 256.4 F 19.4  
24:00 266.9 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.2  
1:00 272.7 F 29.9  
2:00 271.5 F 29.2  
3:00 271.6 F 29.8  
4:00 272.5 F 29.7

Min 271.1 F 29.1  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.6 F 0.0  
5:00 221.3 F 0.0  
10:00 223.3 F 0.0  
15:00 206.4 F 0.0  
20:00 195.2 F 0.0  
25:00 217.4 F 0.0  
30:00 232.5 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:35  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:04:18  
CYCLE COMPLETE

10/24/2024 02:34 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_