
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001268
12:19 22/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 3:05
STERILIZING 4:17
136.0 C 220kPa 4:17
136.2 C 222kPa 4:47
136.2 C 223kPa 5:17
136.0 C 222kPa 5:47
136.2 C 223kPa 6:17
136.2 C 222kPa 6:47
136.3 C 224kPa 7:17
136.4 C 225kPa 7:47
136.4 C 225kPa 7:47
VENTING 7:47
AIR DRYING 8:37
CYCLE COMPLETE 12:49

Digital Signature #
D2A0380E67B3FC4D
