
Midmark M11 - v1.0.5
Total Cycles: 4173

Name: RITTER2
Sterilizer ID: V2313228
12/02/2024 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.0 F 0.0
2:00 96.0 F 0.0
4:00 99.5 F 0.0
6:00 142.3 F 0.5
8:00 162.9 F 2.0
10:00 181.8 F 4.9
12:00 198.9 F 8.1
14:00 210.6 F 11.7
16:00 222.2 F 6.2
18:00 229.0 F 10.7
20:00 233.4 F 8.1
22:00 241.3 F 12.5
24:00 239.8 F 10.7
26:00 251.2 F 16.3
28:00 261.2 F 22.1
30:00 270.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.5
1:00 272.5 F 29.6
2:00 272.7 F 30.0
3:00 271.7 F 29.6
4:00 272.6 F 29.7

Min 271.6 F 29.4
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.5 F 0.0
5:00 210.7 F 0.0
10:00 211.4 F 0.0
15:00 182.1 F 0.0
20:00 175.5 F 0.0
25:00 177.3 F 0.0
30:00 180.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 31:41
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:10:24
CYCLE COMPLETE

12/02/2024 02:14 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____