
Midmark M11 - v1.0.0
Total Cycles: 765

Name: RITTER 1
Sterilizer ID: V2178198
04/01/2021 09:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.7 F 0.0
2:00 195.6 F 2.7
4:00 203.9 F 6.8
6:00 218.2 F 13.2
8:00 223.5 F 9.1
10:00 237.7 F 9.2
12:00 243.5 F 12.9
14:00 242.7 F 11.4
16:00 255.7 F 18.2
18:00 267.2 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.6
1:00 271.5 F 28.7
2:00 272.1 F 29.0
3:00 272.3 F 28.8
4:00 272.0 F 28.7

Min 271.0 F 28.5
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.7 F 0.0
5:00 235.4 F 0.0
10:00 225.9 F 0.0
15:00 225.6 F 0.0
20:00 223.1 F 0.0
25:00 220.0 F 0.0
30:00 210.9 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 19:50
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

04/01/2021 10:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____