
Midmark M11 - v1.0.0
Total Cycles: 1577

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2022 06:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.6 F 0.0
2:00 158.8 F 2.9
4:00 179.0 F 7.3
6:00 203.3 F 12.5
8:00 218.8 F 5.6
10:00 231.8 F 10.6
12:00 232.5 F 7.2
14:00 245.3 F 13.3
16:00 244.3 F 12.1
18:00 257.7 F 19.1
20:00 269.6 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.6 F 28.5
2:00 271.6 F 28.0
3:00 272.8 F 28.6
4:00 271.6 F 27.9

Min 271.4 F 27.8
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 225.8 F 0.0
10:00 217.6 F 0.0
15:00 215.3 F 0.0
20:00 214.4 F 0.0
25:00 215.6 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 20:57
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

01/31/2022 07:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____