
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002171
7:47 28/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:01
STERILIZING 3:57
136.0 C 220kPa 3:57
136.4 C 224kPa 4:27
136.1 C 222kPa 4:57
136.3 C 223kPa 5:27
136.3 C 222kPa 5:57
136.3 C 223kPa 6:27
136.2 C 222kPa 6:57
136.2 C 221kPa 7:27
136.2 C 221kPa 7:27
VENTING 7:27
AIR DRYING 8:22
CYCLE COMPLETE 38:23

Digital Signature #
704F2B07BBC20383
