
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000170
12:33 18/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:08
STERILIZING 4:10
136.4 C 223kPa 4:10
136.3 C 222kPa 4:40
135.9 C 220kPa 5:10
136.3 C 224kPa 5:40
136.1 C 221kPa 6:10
136.4 C 221kPa 6:40
136.4 C 224kPa 7:10
136.4 C 221kPa 7:40
136.4 C 221kPa 7:40
VENTING 7:40
AIR DRYING 8:30
CYCLE COMPLETE 8:37

Digital Signature #
B82E55C68C85E955
