

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2943

Name: RITTER 1  
Sterilizer ID: V2178198  
02/27/2023 08:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.3 F 0.0  
2:00 76.3 F 0.0  
4:00 93.8 F 0.0  
6:00 119.4 F 0.7  
8:00 150.2 F 2.6  
10:00 173.2 F 4.8  
12:00 195.2 F 7.6  
14:00 206.7 F 11.8  
16:00 216.9 F 4.6  
18:00 226.6 F 8.4  
20:00 236.9 F 12.5  
22:00 235.3 F 9.0  
24:00 241.1 F 10.7  
26:00 245.3 F 12.7  
28:00 256.8 F 18.8  
30:00 267.2 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.0 F 28.3  
2:00 272.5 F 28.8  
3:00 271.8 F 28.5  
4:00 271.8 F 28.3

Min 271.4 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.8 F 0.0  
5:00 225.5 F 0.0  
10:00 207.1 F 0.0  
15:00 195.7 F 0.0  
20:00 194.2 F 0.0  
25:00 228.5 F 0.0  
30:00 220.9 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 31:25  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:09:28  
CYCLE COMPLETE

02/27/2023 09:36 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_