
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007834
14:01 06/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:03
STERILIZING 3:04
136.4 C 220kPa 3:04
136.2 C 220kPa 3:34
136.3 C 223kPa 4:04
136.4 C 223kPa 4:34
135.9 C 221kPa 5:04
136.4 C 224kPa 5:34
136.3 C 225kPa 6:04
136.3 C 221kPa 6:34
136.3 C 221kPa 6:34
VENTING 6:34
AIR DRYING 7:25
CYCLE COMPLETE 33:32

Digital Signature #
0EDE23DAF11D4D0A
