
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002024
11:53 12/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:48
PRESSURIZING 1:54
STERILIZING 2:52
136.0 C 220kPa 2:52
136.3 C 223kPa 3:22
136.1 C 224kPa 3:52
136.4 C 227kPa 4:22
136.4 C 225kPa 4:52
136.2 C 225kPa 5:22
136.0 C 223kPa 5:52
136.4 C 227kPa 6:22
136.4 C 227kPa 6:22
VENTING 6:22
AIR DRYING 7:14
CYCLE COMPLETE 37:14

Digital Signature #
AD98680FB6F48BDC
