

-----  
Midmark M11 - v1.0.5  
Total Cycles: 419

Name: RITTER2  
Sterilizer ID: V2313228  
10/06/2021 07:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.7 F 0.0  
2:00 73.2 F 0.0  
4:00 75.7 F 0.0  
6:00 96.8 F 0.2  
8:00 186.6 F 0.6  
10:00 191.5 F 4.0  
12:00 194.2 F 7.5  
14:00 207.1 F 11.3  
16:00 228.6 F 6.2  
18:00 232.5 F 11.4  
20:00 233.9 F 8.4  
22:00 241.4 F 11.7  
24:00 244.3 F 12.7  
26:00 256.1 F 19.0  
28:00 266.4 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.7  
1:00 271.7 F 29.8  
2:00 271.4 F 29.9  
3:00 271.1 F 29.8  
4:00 271.5 F 29.7

Min 271.1 F 29.6  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
5:00 230.3 F 0.0  
10:00 209.2 F 0.0  
15:00 195.1 F 0.0  
20:00 189.7 F 0.0  
25:00 179.0 F 0.0  
30:00 168.8 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 29:57  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:08:07  
CYCLE COMPLETE

10/06/2021 08:34 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01