
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000157
10:51 15/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:36
PRESSURIZING 4:36
STERILIZING 5:48
136.0 C 220kPa 5:48
136.4 C 225kPa 6:18
136.3 C 224kPa 6:48
136.4 C 225kPa 7:18
136.3 C 224kPa 7:48
136.4 C 225kPa 8:18
136.2 C 224kPa 8:48
136.3 C 225kPa 9:18
136.3 C 225kPa 9:18
VENTING 9:18
AIR DRYING 10:20
CYCLE COMPLETE 10:33

Digital Signature #
F11D64CBE9EF3572
