
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008483
8:28 28/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:27
PRESSURIZING 1:31
STERILIZING 2:31
136.4 C 220kPa 2:31
136.3 C 222kPa 3:01
136.3 C 221kPa 3:31
136.3 C 219kPa 4:01
136.4 C 223kPa 4:31
136.3 C 219kPa 5:01
136.3 C 220kPa 5:31
136.3 C 222kPa 6:01
136.3 C 222kPa 6:01
VENTING 6:02
AIR DRYING 6:51
CYCLE COMPLETE 10:09

Digital Signature #
0F1ACA56A0B882E0
