
Midmark M11 - v1.0.0
Total Cycles: 4361

Name: RITTER 1
Sterilizer ID: V2178198
03/06/2024 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 84.5 F 0.0
2:00 85.1 F 0.5
4:00 91.4 F 1.9
6:00 117.7 F 4.1
8:00 144.6 F 6.7
10:00 164.0 F 9.7
12:00 188.2 F 13.1
14:00 205.6 F 4.7
16:00 215.8 F 8.1
18:00 227.3 F 11.8
20:00 228.2 F 6.9
22:00 238.6 F 11.6
24:00 236.3 F 9.0
26:00 248.3 F 14.6
28:00 258.5 F 20.1
30:00 268.0 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.6
1:00 272.3 F 29.3
2:00 271.6 F 28.6
3:00 271.5 F 29.0
4:00 272.6 F 29.8

Min 271.0 F 28.4
Max 273.3 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.1 F 0.0
5:00 228.2 F 0.0
10:00 240.1 F 0.0
15:00 228.5 F 0.0
20:00 227.0 F 0.0
25:00 228.3 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:26
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:10:05
CYCLE COMPLETE

03/06/2024 05:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____