
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001519
9:23 10/21/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:11
PRESSURIZING 6:56
STERILIZING 8:16
136.0 C 220kPa 8:16
136.5 C 218kPa 8:46
136.2 C 221kPa 9:16
136.4 C 220kPa 9:46
136.4 C 220kPa 10:16
136.0 C 218kPa 10:46
136.4 C 223kPa 11:16
136.3 C 221kPa 11:46
VENTING 11:46
AIR DRYING 12:49
CYCLE COMPLETE 28:36

Digital Signature #
639E9D5C01898A16
