

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3567

Name: RITTER2  
Sterilizer ID: V2313228  
04/29/2024 09:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.4 F 0.0  
2:00 73.2 F 0.0  
4:00 106.2 F 0.3  
6:00 124.7 F 2.3  
8:00 144.9 F 5.1  
10:00 166.9 F 8.1  
12:00 184.9 F 10.8  
14:00 202.0 F 6.1  
16:00 212.8 F 5.8  
18:00 222.6 F 9.1  
20:00 232.4 F 12.9  
22:00 231.7 F 8.5  
24:00 242.5 F 13.3  
26:00 241.2 F 11.4  
28:00 252.8 F 17.2  
30:00 263.0 F 23.3  
32:00 272.7 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.9  
1:00 272.9 F 29.8  
2:00 271.9 F 28.9  
3:00 272.2 F 29.6  
4:00 272.8 F 29.5  
  
Min 271.5 F 28.8  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 204.5 F 0.0  
10:00 216.7 F 0.0  
15:00 214.4 F 0.0  
20:00 200.0 F 0.0  
25:00 198.9 F 0.0  
30:00 220.6 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 32:26  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:10:38  
CYCLE COMPLETE

04/29/2024 10:23 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_