

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2677

Name: RITTER 1  
Sterilizer ID: V2178198  
12/14/2022 02:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.9 F 0.0  
2:00 193.4 F 4.1  
4:00 199.1 F 8.9  
6:00 218.0 F 7.7  
8:00 229.4 F 9.2  
10:00 235.5 F 9.2  
12:00 243.5 F 12.7  
14:00 242.7 F 11.6  
16:00 256.3 F 18.7  
18:00 268.3 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.9  
2:00 271.6 F 28.4  
3:00 272.5 F 29.0  
4:00 271.8 F 28.3  
  
Min 271.5 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 230.2 F 0.0  
10:00 241.2 F 0.0  
15:00 235.3 F 0.0  
20:00 230.1 F 0.0  
25:00 229.3 F 0.0  
30:00 231.6 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 19:10  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:41  
CYCLE COMPLETE

12/14/2022 03:54 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_