
Midmark M11 - v1.0.0
Total Cycles: 3649

Name: RITTER 1
Sterilizer ID: V2178198
08/14/2023 06:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.4 F 0.0
2:00 174.4 F 3.3
4:00 194.3 F 7.9
6:00 214.6 F 13.2
8:00 224.0 F 7.3
10:00 236.1 F 12.3
12:00 236.2 F 9.5
14:00 238.2 F 9.3
16:00 248.4 F 14.3
18:00 260.3 F 20.9
20:00 271.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 271.6 F 28.1
2:00 271.8 F 28.2
3:00 272.2 F 28.4
4:00 272.5 F 28.7

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 238.5 F 0.0
10:00 230.4 F 0.0
15:00 227.0 F 0.0
20:00 225.0 F 0.0
25:00 220.5 F 0.0
30:00 207.7 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:36
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:59:49
CYCLE COMPLETE

08/14/2023 07:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____