

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001546  
13:13 24/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:29  
STERILIZING 3:31  
136.5 C 227kPa 3:31  
136.1 C 224kPa 4:01  
136.4 C 227kPa 4:31  
136.3 C 225kPa 5:01  
136.3 C 225kPa 5:31  
136.4 C 226kPa 6:01  
136.4 C 227kPa 6:31  
136.4 C 226kPa 7:01  
136.4 C 226kPa 7:01  
VENTING 7:01  
AIR DRYING 7:54  
CYCLE COMPLETE 10:51  
-----

Digital Signature #  
31D137877401EFA4  
-----