
Midmark M11 - v1.0.5
Total Cycles: 1515

Name: RITTER2
Sterilizer ID: V2313228
09/21/2022 01:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.4 F 0.0
2:00 178.4 F 3.6
4:00 187.8 F 8.2
6:00 209.2 F 13.5
8:00 222.1 F 8.9
10:00 236.1 F 10.4
12:00 238.5 F 11.9
14:00 238.7 F 10.5
16:00 251.5 F 17.0
18:00 262.8 F 23.5
20:00 273.0 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 271.8 F 30.1
2:00 272.7 F 30.1
3:00 272.3 F 29.8
4:00 271.5 F 29.4

Min 271.0 F 29.2
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.4 F 0.0
5:00 218.7 F 0.0
10:00 232.1 F 0.0
15:00 228.4 F 0.0
20:00 226.4 F 0.0
25:00 226.7 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:26
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:16
CYCLE COMPLETE

09/21/2022 02:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____