
Midmark M11 - v1.0.5
Total Cycles: 2130

Name: RITTER2
Sterilizer ID: V2313228
03/08/2023 04:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.6 F 0.0
2:00 188.6 F 3.4
4:00 203.4 F 8.5
6:00 217.5 F 10.0
8:00 230.5 F 9.2
10:00 237.0 F 10.0
12:00 244.2 F 13.2
14:00 244.6 F 12.8
16:00 257.6 F 19.7
18:00 269.1 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.6 F 29.5
2:00 271.8 F 28.7
3:00 271.8 F 29.3
4:00 272.4 F 29.4

Min 271.2 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 215.1 F 0.0
10:00 222.3 F 0.0
15:00 224.3 F 0.0
20:00 223.4 F 0.0
25:00 223.7 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 2:14
HEATING: 19:06
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

03/08/2023 05:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____