
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001831
13:43 21/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:45
STERILIZING 4:56
136.4 C 220kPa 4:56
136.3 C 224kPa 5:27
136.4 C 225kPa 5:57
136.2 C 220kPa 6:27
136.3 C 221kPa 6:57
136.2 C 222kPa 7:27
136.4 C 225kPa 7:57
136.2 C 222kPa 8:27
136.2 C 222kPa 8:27
VENTING 8:27
AIR DRYING 9:27
CYCLE COMPLETE 9:33

Digital Signature #
9E20FEB5E4C5D1F7
