
Midmark M11 - v1.0.5
Total Cycles: 4038

Name: RITTER2
Sterilizer ID: V2313228
10/17/2024 11:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.7 F 0.0
2:00 153.0 F 2.2
4:00 166.0 F 5.8
6:00 193.4 F 10.3
8:00 214.2 F 5.5
10:00 223.6 F 10.3
12:00 229.6 F 8.2
14:00 241.0 F 13.1
16:00 241.3 F 11.9
18:00 254.2 F 18.4
20:00 265.4 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.7 F 30.3
2:00 271.9 F 29.6
3:00 271.8 F 30.0
4:00 272.8 F 30.2

Min 271.3 F 29.4
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 183.3 F 0.0
10:00 208.2 F 0.0
15:00 207.2 F 0.0
20:00 203.6 F 0.0
25:00 202.3 F 0.0
30:00 202.7 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:00:36
CYCLE COMPLETE

10/17/2024 12:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____