

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2975

Name: RITTER 1  
Sterilizer ID: V2178198  
03/06/2023 01:04 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.5 F 0.0  
2:00 168.2 F 3.4  
4:00 187.8 F 7.6  
6:00 212.7 F 12.7  
8:00 223.9 F 7.5  
10:00 235.1 F 12.8  
12:00 237.2 F 9.8  
14:00 235.7 F 8.3  
16:00 248.7 F 14.6  
18:00 260.8 F 21.5  
20:00 271.8 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.7 F 29.3  
2:00 271.6 F 28.5  
3:00 272.7 F 29.5  
4:00 272.2 F 28.9

Min 271.3 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.7 F 0.0  
5:00 240.8 F 0.0  
10:00 228.0 F 0.0  
15:00 229.3 F 0.0  
20:00 228.9 F 0.0  
25:00 230.2 F 0.0  
30:00 229.6 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 20:36  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:50  
CYCLE COMPLETE

03/06/2023 02:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_