
Midmark M11 - v1.0.5
Total Cycles: 1846

Name: RITTER2
Sterilizer ID: V2313228
12/15/2022 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.0 F 0.0
2:00 189.3 F 3.8
4:00 199.3 F 9.5
6:00 216.5 F 6.6
8:00 231.7 F 11.1
10:00 235.1 F 8.8
12:00 238.2 F 9.9
14:00 249.0 F 14.8
16:00 261.4 F 21.9
18:00 272.5 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 29.5
1:00 272.5 F 29.9
2:00 272.4 F 29.9
3:00 272.3 F 29.8
4:00 272.4 F 29.9

Min 272.0 F 29.4
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.0 F 0.0
5:00 217.5 F 0.0
10:00 241.6 F 0.0
15:00 237.8 F 0.0
20:00 233.8 F 0.0
25:00 223.3 F 0.0
30:00 222.4 F 0.0

Drying Complete

FILLING: 2:28
HEATING: 19:05
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

12/15/2022 12:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____