

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001431  
9:39 31/03/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:29  
PRESSURIZING 2:32  
STERILIZING 3:33  
136.0 C 220kPa 3:33  
136.4 C 224kPa 4:03  
136.0 C 221kPa 4:33  
136.4 C 225kPa 5:03  
136.3 C 221kPa 5:33  
136.2 C 224kPa 6:03  
136.4 C 224kPa 6:33  
136.4 C 225kPa 7:03  
136.4 C 225kPa 7:03  
VENTING 7:04  
AIR DRYING 7:57  
CYCLE COMPLETE 8:27  
-----

Digital Signature #  
594F2330FCE9ED78  
-----