
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000940
9:36 09/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:44
STERILIZING 3:47
136.4 C 223kPa 3:47
136.3 C 223kPa 4:17
136.3 C 220kPa 4:47
136.3 C 223kPa 5:17
136.1 C 223kPa 5:47
136.4 C 224kPa 6:17
136.4 C 224kPa 6:47
136.1 C 217kPa 7:17
136.1 C 217kPa 7:17
VENTING 7:17
AIR DRYING 8:09
CYCLE COMPLETE 8:27

Digital Signature #
A0154AAACCEA342B
