

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4873

Name: RITTER 1  
Sterilizer ID: V2178198  
09/06/2024 11:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.1 F 0.0  
2:00 180.5 F 0.6  
4:00 209.0 F 4.2  
6:00 214.8 F 8.9  
8:00 231.0 F 11.2  
10:00 233.6 F 9.3  
12:00 238.4 F 9.6  
14:00 244.9 F 13.1  
16:00 244.2 F 12.2  
18:00 257.0 F 18.8  
20:00 268.0 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.7 F 28.9  
2:00 271.7 F 28.1  
3:00 272.4 F 28.9  
4:00 272.1 F 28.5  
  
Min 271.4 F 28.0  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.1 F 0.0  
5:00 229.4 F 0.0  
10:00 227.3 F 0.0  
15:00 229.8 F 0.0  
20:00 230.6 F 0.0  
25:00 227.5 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 21:19  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:00:18  
CYCLE COMPLETE

09/06/2024 12:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_