
Midmark M11 - v1.0.0
Total Cycles: 5062

Name: RITTER 1
Sterilizer ID: V2178198
11/06/2024 09:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.7 F 0.0
2:00 77.1 F 0.0
4:00 90.7 F 0.0
6:00 123.4 F 0.1
8:00 152.4 F 0.3
10:00 182.1 F 0.6
12:00 198.7 F 3.4
14:00 212.5 F 6.4
16:00 220.5 F 9.9
18:00 230.8 F 12.2
20:00 231.2 F 8.6
22:00 242.1 F 13.4
24:00 240.3 F 10.7
26:00 235.8 F 8.6
28:00 250.2 F 15.2
30:00 261.3 F 21.5
32:00 271.8 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.7 F 28.9
2:00 271.9 F 28.3
3:00 271.8 F 28.4
4:00 272.7 F 28.9

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 197.8 F 0.0
10:00 189.2 F 0.0
15:00 187.7 F 0.0
20:00 188.6 F 0.0
25:00 222.1 F 0.0
30:00 225.8 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 32:39
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:10:45
CYCLE COMPLETE

11/06/2024 10:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____