
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003599
9:33 04/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:03
STERILIZING 2:58
136.0 C 220kPa 2:58
136.3 C 224kPa 3:28
136.4 C 220kPa 3:58
136.4 C 225kPa 4:28
136.4 C 223kPa 4:58
136.4 C 223kPa 5:28
136.4 C 223kPa 5:58
136.4 C 223kPa 6:28
136.4 C 223kPa 6:28
VENTING 6:28
AIR DRYING 7:17
CYCLE COMPLETE 10:07

Digital Signature #
70FCE7FC0279070D
