
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000456
14:11 01/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:06
STERILIZING 4:08
136.0 C 228kPa 4:08
136.4 C 233kPa 4:38
136.4 C 230kPa 5:08
136.3 C 232kPa 5:38
136.3 C 233kPa 6:08
136.2 C 232kPa 6:38
136.4 C 233kPa 7:08
136.4 C 233kPa 7:38
136.4 C 233kPa 7:38
VENTING 7:39
AIR DRYING 8:34
CYCLE COMPLETE 8:49

Digital Signature #
3265F554FD9F2D26
