
Midmark M11 - v1.0.0
Total Cycles: 4559

Name: RITTER 1
Sterilizer ID: V2178198
05/15/2024 02:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.9 F 0.0
2:00 171.7 F 3.3
4:00 183.7 F 7.4
6:00 206.2 F 12.5
8:00 218.4 F 6.0
10:00 229.7 F 10.4
12:00 232.5 F 7.7
14:00 242.3 F 12.6
16:00 241.0 F 11.0
18:00 253.5 F 17.3
20:00 264.7 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.4 F 29.2
2:00 272.2 F 28.7
3:00 271.7 F 28.7
4:00 272.5 F 29.1

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 214.7 F 0.0
10:00 226.8 F 0.0
15:00 232.1 F 0.0
20:00 230.1 F 0.0
25:00 222.6 F 0.0
30:00 220.5 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

05/15/2024 03:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____