
Midmark M11 - v1.0.5
Total Cycles: 968

Name: RITTER2
Sterilizer ID: V2313228
04/07/2022 10:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.0 F 0.0
2:00 140.9 F 0.1
4:00 168.0 F 0.8
6:00 193.4 F 4.7
8:00 212.4 F 9.5
10:00 225.0 F 8.0
12:00 233.9 F 10.6
14:00 233.7 F 8.3
16:00 240.5 F 10.7
18:00 247.7 F 14.5
20:00 260.2 F 21.5
22:00 271.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.5 F 29.6
2:00 271.5 F 28.9
3:00 272.5 F 29.7
4:00 271.8 F 29.0

Min 271.4 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 199.1 F 0.0
10:00 203.4 F 0.0
15:00 209.1 F 0.0
20:00 218.0 F 0.0
25:00 222.1 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:40
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:02
CYCLE COMPLETE

04/07/2022 11:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____