
Midmark M11 - v1.0.5
Total Cycles: 3464

Name: RITTER2
Sterilizer ID: V2313228
03/21/2024 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.0 F 0.0
2:00 110.9 F 0.0
4:00 123.9 F 0.9
6:00 148.6 F 3.4
8:00 167.4 F 6.7
10:00 202.4 F 10.4
12:00 221.0 F 7.2
14:00 222.3 F 7.1
16:00 231.9 F 11.1
18:00 232.3 F 7.4
20:00 242.3 F 12.6
22:00 240.1 F 10.9
24:00 251.6 F 16.7
26:00 262.2 F 22.8
28:00 272.0 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.7 F 29.5
2:00 272.2 F 29.9
3:00 271.4 F 29.3
4:00 272.9 F 29.7

Min 271.4 F 29.2
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.3 F 0.0
5:00 219.4 F 0.0
10:00 190.1 F 0.0
15:00 194.1 F 0.0
20:00 184.4 F 0.0
25:00 201.6 F 0.0
30:00 210.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 29:16
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:07:47
CYCLE COMPLETE

03/21/2024 11:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____