

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001583  
13:09 11/15/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:55  
PRESSURIZING 6:36  
STERILIZING 8:05  
136.0 C 220kPa 8:05  
135.5 C 217kPa 8:35  
136.3 C 222kPa 9:05  
136.4 C 223kPa 9:35  
136.0 C 220kPa 10:05  
136.3 C 222kPa 10:35  
136.3 C 222kPa 11:05  
136.4 C 223kPa 11:35  
VENTING 11:35  
AIR DRYING 12:34  
CYCLE COMPLETE 42:34  
-----

Digital Signature #  
60A28A77DDEB93EB  
-----